

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

Rugged Tablet

Name and address of the applicant
Nom et adresse du demandeur

ADLINK Technology Inc.
9F, No.166 Jian Yi Road, Zhonghe District New Taipei City, 235
Taiwan

Name and address of the manufacturer
Nom et adresse du fabricant

ADLINK Technology Inc.
9F, No.166 Jian Yi Road, Zhonghe District New Taipei City, 235
Taiwan

Name and address of the factory
Nom et adresse de l'usine

ADLINK Technology Inc. JIAN YI FACTORY
4F and 15F, NO. 166, JIAN YI RD., ZHONGHE DIST., NEW
TAIPEI CITY 235
TAIWAN

Note: When more than one factory, please report on page 2
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2^{ème} page

Additional Information on page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

(Sinpro Electronics Co Ltd, model: HPU32A-105)
Input rating: 100-240 Vac, 47-63 Hz, 0.6-0.4 A
Output rating: 12Vdc/2.5A
Rugged Tablet Input Rating: 12Vdc, 2.5A

Trademark (if any)
Marque de fabrique (si elle existe)

ADLINK Technology Inc.



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

IMT-BT

Model / Type Ref.
Ref. De type

Additionally evaluated to EN 60601-1: 2006/ A1: 2013/ A12:2014

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{ème} page

Additional Information on page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1, IEC 60601-1-
6(ed.3), IEC 60601-1-6(ed.3);am1, IEC 62366(ed.1), IEC
62366(ed.1);am1

As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

160701503 issued on 2016-11-04

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-11-07

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-59186-UL

Factories:

ADLINK Technology (China) Co., Ltd.
300 Fang Chun Rd., Zhangjiang Hi-Tech Park, Pudong New Area, Shanghai, 201203
China

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-11-07

Signature:

Jan-Erik Storgaard